

複写



JAB



試験所 認定証

認定番号 RTL02190

機 関 名 称 : ローム株式会社
品質本部 分析センター

所 在 地 : 京都府京都市右京区西院溝崎町 2 1 番地

貴機関は本協会の下記の基準に適合していることが認められましたので、ここに試験所として認定します。

適 用 基 準 : JIS Q 17025:2018 (ISO/IEC 17025:2017)

認 定 範 囲 : 化学試験 (附属書による。)

事 業 所 : 附属書による。

有 効 期 限 : 2023 年 3 月 31 日

第 6 回改定日 2020 年 9 月 9 日

第 3 回更新日 2019 年 2 月 18 日

初回認定日 2007 年 3 月 27 日

公益財団法人

日本適合性認定協会

理事長

飯塚悦功

飯 塚 悦 功

複写



JAB



認定番号

RTL02190

認定証 附属書

(1/1 頁)

試験所・校正機関の別	試験所
機関名称	ローム株式会社 品質本部 分析センター
機関所在地	京都府京都市右京区西院溝崎町 2 1 番地

1) 試験を実施する事業所

事業所名称	ローム株式会社 品質本部 分析センター	
同 所在地	〒	615-8585
	住所	京都府京都市右京区西院溝崎町 21
恒久的施設で行う試験か、 現地試験かの別	☒ 恒久的施設で行う試験 ☐ 現地試験	

認定範囲

分野	M26 化学試験
分類コード	M26.A8
分類名称	樹脂・ゴム

技術分類コード 及び名称	試験対象項目	試験規格／標準作業手順書
B3.1 蛍光 X 線分析	$25 \text{ mg/kg} \leq \text{Cd} \leq 1000 \text{ mg/kg}$ $50 \text{ mg/kg} \leq \text{Pb} \leq 1000 \text{ mg/kg}$ $100 \text{ mg/kg} \leq \text{Cr} \leq 1000 \text{ mg/kg}$	IEC 62321-3-1 一部変更 (定量法に変更)

公益財団法人
日本適合性認定協会

COPY



JAB



Testing Laboratory
Accreditation
Certificate

Accreditation No. RTL02190

ROHM CO., LTD.
Corporate Quality Headquarters Analysis Center

**21, Saiin Mizosakicho, Ukyo-ku, Kyoto-shi, Kyoto, 615-8585
Japan**

meets the following criteria. On the basis of this, Japan Accreditation Board (JAB) grants accreditation to the said testing laboratory.

Applicable accreditation criteria	: JIS Q 17025:2018 (ISO/IEC 17025:2017)
Scope of accreditation	: Chemical testing (As described in the appendix)
Premises covered by accreditation	: As described in the appendix.
Expiry date of accreditation	: March 31, 2023

Revised (6)	September 9, 2020
Renewed (3)	February 18, 2019
Initial accreditation	March 27, 2007

Y. Iizuka, President

Japan Accreditation Board

COPY

Accreditation No.

RTL02190



JAB



Accreditation Certificate

Appendix

(Page 1/1)

Type of Laboratory	Testing Laboratory
Name of Laboratory	ROHM CO., LTD. Corporate Quality Headquarters Analysis Center
Address	21, Saiin Mizosakicho, Ukyo-ku, Kyoto-shi, Kyoto, 615-8585 Japan

1) Premises on which testing activities are performed

Name of Premises	ROHM CO., LTD. Corporate Quality Headquarters Analysis Lab		
Address of Premises	Postal code	615-8585	
	Address	21, Saiin Mizosaki-cho, Ukyo-ku, Kyoto 615-8585 Japan	
Testing service at permanent facilities or on site testing service		☒ Testing service at permanent facilities ☐ On site testing service	

Scope of Accreditation

FIELD	M26 Chemical Testing
CODE OF CIT*1	M26.A8
NAME OF CIT	Resin, rubber

*1 CIT: Classification of Item to be Tested

*2 TCT: Technical Classification of Test

CODE & NAME OF TCT*2	PROPERTIES MEASURED	TEST METHOD STANDARD / STANDARD OPERATING PROCEDURE
B3.1 X-ray fluorescence Spectrometric analysis	Cd 25 mg/kg to 1000 mg/kg Pb 50 mg/kg to 1000 mg/kg Cr 100 mg/kg to 1000 mg/kg	IEC 62321-3-1 (changed from screening method to quantification method)

Japan Accreditation Board

Issue No. : RTL02190-20200909